




SN# 290575-09 Lamp Specific Acceptance Test

Model: Magnum GO	Result	Steady State	Acceptable	Units
Peak Output Wavelength	Pass	368 ✓	365 ± 5	nm
Full Width Half Maximum (FWHM)	Pass	13 ✓	≤ 15	nm
Full Width at 10% Maximum	Pass	25 ✓	≤ 30	nm
Longest Wavelength at Half Maximum	Pass	375 ✓	≤ 377	nm
Max Excitation Irradiance (347-382 nm)	Pass	3862 ✓	≥ 2000	μW/cm ²



Inspector: VEP
Date of Certification: 6-11-2024

This lamp has been tested to conform to all requirements of:
ASTM E3022 ✓ AITM6-1001 ✓ ISO 3059 ✓ RRES 90061 ✓

C4 Magnum Series, Certificate of Conformance



SN# 290575-09	Switch On	Steady State	Elevated Temp*	Acceptable	Units
Peak Output Wavelength	367	368	370	365 ± 5	nm
Full Width Half Maximum (FWHM)	13	13	13	≤ 15	nm
Full Width at 10% Maximum	25	25	27	≤ 30	nm
Longest Wavelength at Half Maximum	374	375	376	≤ 377	nm
Visible Emission at 15 in	9	8	8	≤ 20	Lux
Visible Emission at 36 in	3	3	3	≤ 5	Lux
Irradiated Area at 15 in (>1000 μW/cm ²)	11	10	9	≥ 5	Inch Diameter
Max Excitation Irradiance (347-382 nm)	4,248	3,862	2,400	≥ 2,000	μW/cm ²
Irradiance at Peak Output Wavelength	312	284	185	-	μW/cm ²
Min Working Distance** (< 5000 μW/cm ²)	14	12	11	-	Inches
Max Working Distance** (> 1200 μW/cm ²)	26	23	21	-	Inches
Body Stabilization Temp (70°F Ambient)	-	116	-	≤ 120	°F
Current Ripple	9	9	9	≤ 35	mA peak-peak

* Tested in a 110 °F Environment, the **maximum ambient temperature where still light meets all requirements**

** C4 Magnum series lights use a single LED emitter, thus beam uniformity does not change with working distance (RRES 90061)

