




SN# 290633-04 Lamp Specific Acceptance Test

Model: Nomad-Eclipse	Result	Steady State	Acceptable	Units
Peak Output Wavelength	Pass	370 ✓	365 ± 5	nm
Full Width Half Maximum (FWHM)	Pass	11 ✓	≤ 15	nm
Full Width at 10% Maximum	Pass	24 ✓	≤ 30	nm
Longest Wavelength at Half Maximum	Pass	375 ✓	≤ 377	nm
Max Excitation Irradiance (347-382 nm)	Pass	2013 ✓	≥ 1200	μW/cm ²



Inspector: VEP
Date of Certification: 9-29-2024

This lamp has been tested to conform to all requirements of:
ASTM E3022 ✓ AITM6-1001 ✓ ISO 3059 ✓ RRES 90061 ✓

Nomad-Eclipse, Certificate of Conformance




SN# 290633-04	Switch On	Steady State	Elevated Temp*	Acceptable	Units
Peak Output Wavelength	369	370	370	365 ± 5	nm
Full Width Half Maximum (FWHM)	11	11	12	≤ 15	nm
Full Width at 10% Maximum	24	24	24	≤ 30	nm
Longest Wavelength at Half Maximum	375	375	375	≤ 377	nm
Visible Emission at 12 in	11	10	8	≤ 20	Lux
Visible Emission at 36 in	2	2	2	≤ 5	Lux
Irradiated Area at 12 in (>1000 μW/cm ²)	12	10	9	≥ 5	Inch Diameter
Max Excitation Irradiance (347-382 nm)	2,516	2,013	1,400	≥ 1200	μW/cm ²
Irradiance at Peak Output Wavelength	215	172	120	-	μW/cm ²
Min Working Distance** (< 5000 μW/cm ²)	7	7	6	-	Inches
Max Working Distance** (> 1200 μW/cm ²)	15	14	13	-	Inches
Handle Stabilization Temp (70°F Ambient)	-	120	-	≤ 120	°F
Current Ripple	26	26	26	≤ 58	mA peak-peak
Typical Battery Discharge Time (18Wh Batt)	-	95	-	-	Minutes


* Tested in a 85 °F Environment, the maximum ambient temperature where still light meets requirements

** Nomad-Eclipse has tightly grouped LED emitters, thus beam uniformity does not change with working distance (RRES 90061)


Glass UV Pass filter
Transmittance:
~0% < 275nm
~85% (Peak) @ 365nm
~0% > 400nm < 675nm
Secondary Peak ~42% @ 775nm
<20% above 825nm
Used in forensics, Testing, chemical analysis.
DIA: 25mm T: 1.5mm WT: .004



BEAM PROFILE AT 12 INCHES



Nomad-Eclipse Typical Emission



Performance over time

