




C4 Magnum Series, Certificate of Conformance



SN# 12071603_magnum_recert3	Switch On	Steady State	Elevated Temp*	Acceptable	Units
Peak Output Wavelength	367	368	369	365 ± 5	nm
Full Width Half Maximum (FWHM)	11	11	13	≤ 15	nm
Full Width at 10% Maximum	24	24	25	≤ 30	nm
Longest Wavelength at Half Maximum	374	375	375	≤ 377	nm
Visible Emission at 15 in	9	8	8	≤ 20	Lux
Visible Emission at 36 in	3	3	3	≤ 5	Lux
Irradiated Area at 15 in (>1000 μW/cm ²)	11	10	9	≥ 5	Inch Diameter
Max Excitation Irradiance (347-382 nm)	5,307	4,245	2,400	≥ 2,000	μW/cm ²
Irradiance at Peak Output Wavelength	427	342	185	-	μW/cm ²
Min Working Distance** (< 5000 μW/cm ²)	15	14	13	-	Inches
Max Working Distance** (> 1200 μW/cm ²)	26	23	21	-	Inches
Body Stabilization Temp (70°F Ambient)	-	116	-	≤ 120	°F
Current Ripple	9	9	9	≤ 35	mA peak-peak

* Tested in a **110° F** Environment, the **maximum ambient temperature where still light meets all requirements**


** C4 Magnum series lights use a single LED emitter, thus beam uniformity does not change with working distance (RRES 90061)





SN# 12071603 Lamp Specific Acceptance Test

Model: Magnum	Result	Steady State	Acceptable	Units
Peak Output Wavelength	Pass	368	365 ± 5	nm
Full Width Half Maximum (FWHM)	Pass	11	≤ 15	nm
Full Width at 10% Maximum	Pass	24	≤ 30	nm
Longest Wavelength at Half Maximum	Pass	375	≤ 377	nm
Max Excitation Irradiance (347-382 nm)	Pass	4245	≥ 2000	$\mu\text{W}/\text{cm}^2$



Inspector: _____

Date of Certification: _____

This lamp has been tested to conform to all requirements of:

ASTM E3022 AITM6-1001 ISO 3059 RRES 90061

Emission Spectrum Measured at Stabilization Temperature with Blue-wave UV-25 S/N 22011404, Certificate Number 22011404-UV-CR2-1124