


SN# 000900 Lamp Specific Acceptance Test



Model: UV-2000, Magnum	Result	Steady State	Acceptable	Units
Peak Output Wavelength	Pass	368	365 ± 5	nm
Full Width Half Maximum (FWHM)	Pass	11	≤ 15	nm
Full Width at 10% Maximum	Pass	22	≤ 30	nm
Longest Wavelength at Half Maximum	Pass	375	≤ 377	nm
Max Excitation Irradiance (347-382 nm)	Pass	3598	≥ 2000	μW/cm ²



Inspector: MHP
 Date of Certification: 1-5-2026

This lamp has been tested to conform to all requirements of:
 ASTM E3022 ✓ AITM6-1001 ✓ ISO 3059 ✓ RRES 90061 ✓

Emission Spectrum Measured at Stabilization Temperature with Blue-wave UV-25 S/N 18120729, Certificate Number 18120729-UV-CR2-0925



C4 Magnum Series, Certificate of Conformance



SN# 000900	Switch On	Steady State	Elevated Temp*	Acceptable	Units
Peak Output Wavelength	367	368	369	365 ± 5	nm
Full Width Half Maximum (FWHM)	11	11	13	≤ 15	nm
Full Width at 10% Maximum	22	22	26	≤ 30	nm
Longest Wavelength at Half Maximum	374	375	376	≤ 377	nm
Visible Emission at 15 in	7	6	6	≤ 20	Lux
Visible Emission at 36 in	3	3	3	≤ 5	Lux
Irradiated Area at 15 in (>1000 μW/cm ²)	11	10	9	≥ 5	Inch Diameter
Max Excitation Irradiance (347-382 nm)	3,777	3,598	2,400	≥ 2,000	μW/cm ²
Irradiance at Peak Output Wavelength	328	312	185	-	μW/cm ²
Min Working Distance** (< 5000 μW/cm ²)	15	14	13	-	Inches
Max Working Distance** (> 1200 μW/cm ²)	26	23	21	-	Inches
Body Stabilization Temp (70°F Ambient)	-	116	-	≤ 120	°F
Current Ripple	9	9	9	≤ 35	mA peak-peak

* Tested in a 110°F Environment, the **maximum ambient temperature where light still meets all requirements**

** C4 Magnum series lights use a single LED emitter, thus beam uniformity does not change with working distance (RRES 90061)

